

MR Imaging System

OASIS Velocity

IHE Integration Statement

FUJIFILM Healthcare Corporation

Tokyo, Japan

E1en-FM0072-01

© FUJIFILM Healthcare Corporation 2020. All rights reserved.

IHE Integration Statement			
Vendor	Product Name	Version	Date
FUJIFILM Healthcare Corporation	OASIS Velocity	V7.0 or later	1-Apr-2020
This product implements all transactions required in the IHE Technical Framework to support the IHE Integration Profiles, Actors, and Options listed below:			
Integration Profiles Implemented	Actors Implemented	Options Implemented	
Scheduled Workflow (SWF)	Acquisition Modality	Broad Worklist Query Patient Based Worklist Query	
Patient Information Reconciliation (PIR)	Acquisition Modality	None	
Consistent Presentation of Images (CPI)	Acquisition Modality	None	
	Print Composer	None	
Basic Security (SEC)	Acquisition Modality	Encryption	
	Print Composer		
	Secure Node		
Audit Trail and Node Authentication (ATNA)	Acquisition Modality	Encryption	
	Print Composer		
	Secure Node		
Consistent Time (CT)	Time Client	None	
Key Image Note (KIN)	Acquisition Modality	None	
Portable Data for Imaging (PDI)	Portable Media Creator	None	
Internet address for vendor's IHE information:			
Links to Standards Conformance Statements for the Implementation			
HL7			
DICOM			
Links to general information on the IHE			
In North America: www.ihe.net/	In Europe: www.ihe-europe.org/	In Japan: www.ihe-j.org/	